





中国认可 国际互认 检测 TESTING CNAS L0604

Test Report No.: QDHL2008007820MD Date: AUG.21,2022 Page: 1 of 2

WMG MEDICAL CO., LTD SEGBAISHI PHASE II FACTORY, NO.1 HONGLIU ROAD, SHENZHEN, CHINA

The following sample(s) was/were submitted and identified by the client as:

Sample Description : MEDICAL NITRILE INSPECTION GLOVES

 Style No.
 : SQKC006

 Lot No.
 : 20220707

Lot Size : NOT PROVIDED

Sample Quantity : 300PCS

Sample Receiving Date : AUG.06,2022

Testing Period : AUG.06,2022 TO AUG.21,2022

<u>Test Requested</u> <u>Result</u>

BS EN 455-1:2000 Medical Gloves for Single Use – Part 1: Requirements and Pass

Testing for Freedom from Holes (Clause 5.1)

Remark: - Unless otherwise stated the results shown in this test report refer only to the sample(s) tested. This document cannot be used for publicity, without prior written approval of the SGS.

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.



SGS SGS QDHL2008007820MD

Jessica Gao Approved Signatory



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

OD

7389502

www.sgsgroup.com.cn e sgs.china@sgs.com







中国认可 国际互认 构制 TESTING CNAS L0604

Test Report No.: QDHL2008007820MD Date: AUG.21,2022 Page: 2 of 2

Test Conducted:

BS EN 455-1:2000 Medical gloves for single use – Part 1: Requirements and testing for freedom from holes

Number of test sample	:	200 PCS
Sample size	:	M
Number of non-conforming gloves		1 PC

Clause	Test Items	Result		
5	Watertightness test for detection of holes			
5.1	Referee testing	Pass (See note 1)		

Note	:	1	Sample quantity: 200 PCS, AQL:1.5, Ac:7, Re:8, Found: 1 PC. The sample selecting amount for this clause is deviated to 200 pcs per size as assessed by SGS.	
------	---	---	---	--

Sample Photo:

Received Sample



SGS authenticate the photo on original report only

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD

7389503